

Journal Pre-proof

Use of a Portable, Non-pneumatic Active Compression Device in treatment of Phlebolympheidema: a TEAYS sub-analysis

Todd Berland, MD, Michael Barfield, MD, Ron Winokur, MD, Sandi Davis, DPT, Vicky Ralph, ANP-BC, Nancy Chatham, ANP-BC, Stanley Rockson, MD, Thomas S. Maldonado, MD

PII: S0890-5096(26)00133-0

DOI: <https://doi.org/10.1016/j.avsg.2026.02.039>

Reference: AVSG 8197

To appear in: *Annals of Vascular Surgery*

Received Date: 23 March 2025

Revised Date: 17 January 2026

Accepted Date: 17 February 2026

Please cite this article as: Berland T, Barfield M, Winokur R, Davis S, Ralph V, Chatham N, Rockson S, Maldonado TS, Use of a Portable, Non-pneumatic Active Compression Device in treatment of Phlebolympheidema: a TEAYS sub-analysis, *Annals of Vascular Surgery* (2026), doi: <https://doi.org/10.1016/j.avsg.2026.02.039>.

This is a PDF of an article that has undergone enhancements after acceptance, such as the addition of a cover page and metadata, and formatting for readability. This version will undergo additional copyediting, typesetting and review before it is published in its final form. As such, this version is no longer the Accepted Manuscript, but it is not yet the definitive Version of Record; we are providing this early version to give early visibility of the article. Please note that Elsevier's sharing policy for the Published Journal Article applies to this version, see: <https://www.elsevier.com/about/policies-and-standards/sharing#4-published-journal-article>. Please also note that, during the production process, errors may be discovered which could affect the content, and all legal disclaimers that apply to the journal pertain.

© 2026 Published by Elsevier Inc.



1 **Use of a Portable, Non-pneumatic Active Compression Device in treatment of**
2 **Phlebolymphe^dema: a TEAYS sub-analysis**

3
4
5 Short title: Non-pneumatic vs pneumatic =compression for Phlebolymphe^dema
6

7 Todd Berland MD1, Michael Barfield MD2, Ron Winokur MD3, Sandi Davis DPT4, Vicky
8 Ralph ANP-BC5, Nancy Chatham ANP-BC6, Stanley Rockson MD7, Thomas S
9 Maldonado MD1

10
11 1Vascular Surgery, New York University, NY, USA

12 2 University Surgical Associates, University of Tennessee Health
13 Science Center, Nashville, TN, USA

14 3Vascular and Interventional Radiology, Weill Cornell Medical Center, NY, NY, USA

15 4Davis Care Physical Therapy, NY, NY, USA

16 5Physical Medicine and Rehabilitation, University of Colorado, Denver, CO, USA

17 6 St. Johns Regional Wound Clinic, Hospital Sisters Health Center, Springfield, IL, USA

18 7Cardiovascular Medicine, Stanford University, Stanford, California, USA
19

20 Corresponding Author: Thomas S Maldonado, MD

21 Division of Vascular Surgery, New York University Langone Health, New York, NY
22 10016

23 Email: Thomas.maldonado@nyulangone.org
24

25 FUNDING This work was funded by Koya Medical. Koya medical was not involved in manuscript writing
26

27 Keywords:

28 NPCD, Non-Pneumatic Compression, pneumatic compression, Dayspring,
29 Lymphedema Treatment, Lower Extremity Lymphedema, Phlebolymphe^dema
30

Use of a Portable, Non-pneumatic Active Compression Device in treatment of Phlebolymphe^dema: a TEAYS sub-analysis

33

Abstract

Objective

Non-pneumatic compression devices (NPCDs) have demonstrated their clinical efficacy and safety in treating lymphedema (LED) in multiple studies, including two recent multi-centered, randomized head-to-head comparative studies with advanced pneumatic compression devices (APCD). In the most recent study, TEAYS (ClinicalTrials.gov Identifier: NCT05507346), NPCDs demonstrated better clinical utility as well as greater efficacy and adherence than APCDs in the treatment of lower extremity swelling. This current sub analysis of TEAYS focuses on the outcomes for patients whose secondary lymphedema is associated with underlying venous etiology or phlebolymphe^dema (PLED).

Methods

This trial was a randomized, crossover head-to-head study was performed across nine sites in the US in 2023. Patients were subjected to an initial 4-week washout period and then randomized to either the NPCD or a commercially available APCD. Patients used the randomly assigned initial device for 90 days followed by a second 4-week washout period before a 90-day use of the second device. The current study focuses specifically on the sub-analysis of the cohort of PLED patients. Primary efficacy outcomes assessed in this study included change in affected limb volume between baseline (day 0) and end of treatment (day 90), change in Lymphedema Quality of Life Questionnaire (LYMQOL), and treatment adherence.

Results

Analysis included a total of 71 patients with lower extremity lymphedema; 35 of whom were diagnosed with PLED and this subset comprises the study cohort for the current study. In the PLED cohort 13 (37%) were male, Average BMI was 36.2 +/- 1.68, and 19 had bilateral limbs affected (54%). Most patients had clinical stage II lymphedema: I (n=6), II (n=20) and III (n=9). These PLED patients achieved statistically greater mean limb volume reduction (424.49±100.9mL) while on NPCD vs (50.8± 112.1mL) for APCD ($p=.0085$). NPCD also showed significantly better improvement in overall Quality of Life (1.39±0.39) vs. APCD (0.18±0.29); ($p=0.01$). Statistically significant improvement in adherence was also observed while on NPCD 81% vs APCD 49% ($p \leq .001$). No device-related adverse events were reported.

Conclusion

The NPCD is a clinically effective treatment for decreasing limb volume in patients with lower extremity LED. The NPCD was more effective than an APCD and resulted in

69 superior limb volume decrease, greater improved quality of life, adherence, mobility,
70 and patient satisfaction. The outcomes for the subset of patients diagnosed with PLED
71 corroborates the improvements seen in the overall LED study patient population
72 previously reported. Additionally, results suggest that PLED patients may potentially
73 benefit even more from NPCD than non-LED patients.

74

75 **Introduction**

76 Secondary lymphedema (LED) can result as a sequela from a variety of insults to the
77 lymphatic system, including trauma, surgery, cancer and cancer-related treatments,
78 most notably radiation or oncologic surgeries. When secondary lymphedema is the
79 associated with chronic venous hypertension, the condition is called
80 Phlebolymphe­dema (PLED). A large retrospective study found this type of disease as a
81 leading cause of lymphedema in the US.¹⁻⁷

82 The chronic and progressive nature of PLED requires a commitment to lifelong
83 management, which include use of compression garments as well as compression
84 devices in the home setting. Most recently, a mobile non-pneumatic compression
85 device (NPCD) has demonstrated its clinical efficacy and safety in treating LED. The
86 NPCD technology uses shape memory alloy (nickel/titanium) actuators in its garment,
87 which contract and relax to achieve sequential gradient compression in a distal to
88 proximal manner when specified and energized by the controller. In use, the NPCD
89 controller is battery powered, and is designed to allow the patient to retain mobility while
90 performing their activities for daily living versus immobilizing the patient in a supine
91 position during a pneumatic compression treatment. In two separate multi-centered,
92 randomized head-to-head comparative studies with advanced pneumatic compression
93 devices (APCD), which is considered current standard of care, NPCD showed superior
94 outcomes. In the most recent study (Treatment Effectiveness of a Non-Pneumatic
95 Compression Device versus an Advanced Pneumatic Compression Device for Lower
96 Extremity Lymphedema Swelling: TEAYS study), NPCDs demonstrated better clinical
97 utility as well as greater efficacy and adherence compared to APCDs.⁸⁻¹¹ The current
98 study presented here evaluates key outcomes for the subset of patients from the
99 TEAYS study who are diagnosed with PLED.

100 **Methods**

101 The study design for the TEAYS study, a prospective, multicenter, randomized, single,
102 crossover clinical trial conducted across nine study sites in the United States, has been
103 previously described¹¹. Eligible patients with a confirmed diagnosis of primary or
104 secondary unilateral or bilateral lower extremity lymphedema were included. The
105 current study focuses specifically on the sub-analysis of the cohort of PLED patients.

106 Primary efficacy outcomes assessed in this study included change in affected limb
107 volume between baseline (day 0) and end of treatment (day 90), change in

108 Lymphedema Quality of Life Questionnaire (LYMQOL), and treatment adherence.
109 Calculation of limb volume by circumference measure was performed by a trained
110 therapist using a calibrated tape measure. Measurements were taken every 4 cm, and
111 the volume of a truncated cone is calculated according to the Kuhnke formula, summing
112 the eight neighboring circumference measures. Measurements were performed for all
113 affected limbs, regardless of whether lymphedema was unilateral or bilateral.

114 For the QoL assessment, limb specific LYMQOL survey was used (Appendix). The
115 LYMQOL is a 20-item clinically validated disease-specific survey tool, that was
116 administered at days 0 and 90 for each device treatment period. The survey assesses
117 the effects of lymphedema on QOL through both an overall score (scored 1-10) and four
118 sub scores: symptoms (pain, swelling, numbness), body image and appearance,
119 function (activities of daily living; e.g., eating, writing, and dressing), and mood (e.g.,
120 sleep disruption, depression, and irritability). The subdomains are scored from 1 (not at
121 all) to 4 (a lot). The total score is calculated by summing all scores and dividing by the
122 total number of items. The domain-specific sub scores reflect improvement as a lower
123 score, and the overall QOL score reflects improvement by a higher score. Changes
124 from days 0 to day 90 for the total score and each sub score were calculated.

125 Treatment adherence was reported through patient diaries over the 90-day course of
126 treatment for each device. Adherence was calculated as the percentage of reported
127 daily use (minimum of 1 hour) over the treatment period (i.e., patients who used device
128 for the entire 90 days achieved 100% adherence, whereas those who used device
129 every other day reported 50% adherence).

130 Secondary outcomes included safety as measured by device-related adverse events
131 (AEs) (e.g., pressure-induced wounds, allergic reactions to garments, pain from use of
132 device, or thermal burns) throughout the course of study, and a patient survey
133 administered at the end of the study. The survey evaluated the patient's preference for
134 treatment modality as well as their perceived mobility and device portability during
135 treatment and whether they experienced decreased use of their compression garments
136 during each treatment period. Reports on truncal swelling before and after device use
137 were also collected.

138 Additional disease-related health episodes and resource utilization were collected,
139 including episodes of cellulitis, ulceration, hospitalization, lymphedema-related physical
140 therapy visits, and compression stocking use over the past 12 months before device
141 use and during the study duration with each device treatment.

142 Randomization and treatment

143 The study design is depicted in Figure 1. An initial 30-day washout period was
144 established in which no compression devices were used. During this period, patients
145 were allowed to continue their conservative care, which included the use of
146 compression garments, without any physical therapy visits. After this initial 30-day
147 period, each patient was randomized to receive either the NPCD or the APCD treatment

148 for 90 continuous days. At the end of the treatment duration (day 90), another 30-day
149 washout period was established in which no compression device was used, and
150 patients were subsequently crossed over to the alternate device treatment. For each
151 device treatment arm, measurements were collected at day 0 and day 90, except for the
152 patient study survey, which was performed at the end of the study. All patients were
153 trained on how to use the devices and don/doff the respective device accessory
154 garments. Study devices included either the NPCD Dayspring device (Koya Medical,
155 Dallas, TX) or a commercially available APCD (of the 71 patients who completed the
156 study, 2 used an Airos E0652 device (AIROS Medical, Audubon, PA), one used a
157 Lympha Press E0652 device (Lympha Press, Chadds Ford, PA), and the remaining 68
158 patients used the E0652 Flexitouch plus [PG32-G3] device (Tactile Medical, Minneapolis,
159 MN). Patients were instructed to use the assigned device once daily on the study limb
160 for a minimum of 60 minutes. Patients were permitted to continue the use of
161 compression garments and the general duration of use was captured using the patient
162 survey at the end of the study.

163

164 Statistical analysis

165 The software packages used for data analysis for this prospective study were Microsoft
166 Excel (Office Professional Plus 2021, Version 2508; Microsoft Corp, Redmond, WA)
167 and STATA (StataCorp, College Station, TX). Changes in measured outcomes from
168 days 0 to day 90 for both groups and categorical variables were presented as
169 proportions, normally distributed continuous variables presented as mean \pm standard
170 error, and skewed continuous variables presented as median (interquartile range).
171 Assumptions were checked; nonparametric alternatives were considered as needed for
172 skewed distributions. Univariate and multivariable analyses were performed with
173 candidate variables and outcome measures. Statistical significance was tested using a
174 two-sided alpha level of 0.05 and with appropriate multiple testing correction (Bonferroni
175 or Benjamini-Hochberg) approach when needed, with each limb considered a unique
176 observation. A priori sample size calculation was performed based on the primary endpoint of
177 percentage reduction in limb volume (converted from circumference). Assuming an effect size of
178 10% reduction in volume (standard deviation of 12%), with a two-sided $\alpha = 0.05$ and power =
179 80%, we estimated that 40 participants per arm would be required. We enrolled significantly
180 higher than this threshold (plus the cross over effect per arm), exceeding the requirement and
181 thereby ensuring sufficient power to detect the hypothesized difference.

182

183 Study Results

184 Patient demographics

185 A total of 121 patients were screened and 22 failed at screening; 99 patients entered
186 the study. Over the entire study, 24 patients withdrew consent and 4 were lost to follow-
187 up or were missing data. Of the 24 patients who withdrew consent, 3 dropped out of the
188 study before assignment of a treatment device, 6 during the APCD group, and 15 during
189 the NPCD group. All patients had a confirmed diagnosis of lymphedema. Diagnosis of

190 lymphedema was clinical (lymphoscintigraphy was not required; diagnosis of CVD was based
 191 on h/o DVT or documented venous reflux on duplex ultrasound). All patients were on
 192 conservative therapy (including, but not limited to exercise, manual lymphatic drainage,
 193 compression garments, and elevation of limb) before day 0. There were 35 patients
 194 (49%) who were diagnosed with PLED and 36 patients (51%) who were diagnosed with
 195 non-LED. For this subset analysis, the demographics of the PLED subjects compared
 196 to the non-phlebolymphedema (non-LED) subjects are summarized in Table 1.

197 **Primary end points and efficacy**

198 Edema reduction and limb volume

199 In the PLED group, for the NPCD treatment arm, mean limb volume decrease with
 200 standard error of 424.4 ± 100.9 mL ($p = .0011$) and a median of 349 mL was achieved
 201 vs that of 50.8 ± 112.06 mL ($p = .65$) and a median of 7.5 mL for the APCD treatment
 202 arm (Figure 2). Statistical significance for comparing mean limb volume decreases
 203 between the treatment arms was achieved, favoring NPCD ($p = .0085$).

204 In the non-LED group, for the NPCD treatment arm, mean limb volume decrease with
 205 standard error of 317.3 ± 92.42 mL ($p = .0049$) and a median of 210 mL was achieved
 206 vs that of 114.9 ± 78.7 mL ($p = .42$) and a median of 73 mL for the APCD treatment arm
 207 (Figure 2). Statistical significance for comparing mean limb volume decreases between
 208 the treatment arms was achieved, favoring NPCD ($p = .034$).

209 Changes in the foot were monitored by measurements at the metatarsal heads and
 210 midfoot for both treatment groups between day 0 and day 90, and no significant
 211 difference was detected between either group for the PLED or the non-LED subsets
 212 (Figure 3).

213 Quality of Life (LYMQOL)

214 In the PLED group, significant improvement in QoL was achieved for NPCD and but not
 215 for APCD treatment. Overall LYMQOL score improvements of 1.39 ± 0.39 ($p = .015$)
 216 and a median of 1.0 for NPCD vs that of 0.18 ± 0.29 ($p = .53$) and a median of 0.0 for
 217 APCD were achieved. Statistical significance for comparing overall LYMQOL
 218 improvement between the two treatment arms was achieved, favoring NPCD ($p = .01$).
 219 Significant improvement in LYMQOL functional sub-scores were mixed for both
 220 treatment arms. The NPCD treatment arm achieved statistically significant improvement
 221 across all but one sub-score (mood, -0.16 ; $p = .11$), whereas the APCD treatment arm
 222 achieved no statistical significance in any of the sub-scores. Statistical significance for
 223 comparing the LYMQOL functional sub-score improvements between the two treatment
 224 arms was achieved in function, appearance, and symptoms, favoring NPCD, but not in
 225 mood.

226 In the non-LED group, while there were improvements in QoL, they were not
 227 significant for either NPCD or APCD. Overall LYMQOL score improvements of $0.64 \pm$
 228 0.24 ($p = .103$) and a median of 0 for NPCD vs that of 0.16 ± 0.23 ($p = .51$) and a

229 median of 0.0 for APCD were achieved. Refer to Table 2 and Figures 4 and 5 for a
230 summary of the primary outcomes, including LYMQOL. LYMQOL is a validated clinical
231 tool and 1.0 point (the lowest count) in the overall score is considered clinically
232 meaningful.

233 Treatment adherence

234 In the PLED group, treatment adherence was reported as $81\% \pm 4\%$ (with a median of
235 92% for NPCD and $49\% \pm 6.5\%$ with a median of 46% for APCD. Statistical
236 significance was achieved comparing adherence for the two treatment arms, favoring
237 NPCD ($p \leq .001$).

238 In the non-LED group, treatment adherence was reported as $80\% \pm 4.2\%$ with a
239 median of 89% for NPCD and $62\% \pm 5.3\%$ with a median of 69% for APCD. Statistical
240 significance was achieved comparing adherence for the two treatment arms, favoring
241 NPCD ($p = .0024$).

242 Figure 6 contains a graphical representation of these results.

243 Secondary end points and safety

244 No device-related AEs or device-related severe AEs were reported in either the NPCD
245 or the APCD treatment arms. Unrelated to either devices, the following AEs were
246 reported during the course of the study: 2 mild AEs (with a fall on ice and rolled ankle,
247 both resolved); 12 moderate AEs (Mohs surgery, torn calf muscle, allergy to medication,
248 COVID, sciatic leg pain, ankle sprain, implantation of a heart loop recorder, knee
249 injections for pain, cellulitis, allergy to Bactrim; all resolved with medical or surgical
250 intervention; and cancer recurrence managed with ongoing medical intervention); 15
251 moderate severe AEs (cardiac arrhythmia, hospitalization, pacemaker implantation, neck
252 pain/fusion with rod placement, fall with metatarsal break and numbness, torn
253 meniscus, urinary tract infection requiring hospitalization, urinary retention, retinal
254 surgery, cataract surgery, wound vac treatment, and stent placement; all resolved with
255 medical or surgical intervention).

256 No truncal swelling or worsening was reported (compared with baseline) for any
257 patients for either group and for either of the subsets (LED or Non-LED).

258 For the patient survey, which was administered at the end of the study, a majority of the
259 patients (85% in the LED group and 97% in the Non-LED group) reported being
260 active during NPCD treatment (0% for APCD treatment for LED and Non-LED
261 group). Patients also responded on their overall preference, with a majority preferring
262 NPCD treatment as their treatment choice (85% in the LED group and 71% in the
263 Non-LED group) compared to (15% in the LED group and 29% in the Non-LED
264 group) for the APCD treatment.

265 Additionally, patients also reported decreased use of compression stockings in both
266 groups. For the LED subset, 68% of patients on NPCD treatment reported decreased

267 use of compression stockings compared with 12% of patients on APCD treatment
268 reporting decreased use of compression stockings. Similarly, for the non-PLED subset,
269 64% of patients on NPCD treatment reported decreased use of compression stockings
270 compared with 6% of patients on APCD treatment reporting decreased use of
271 compression stockings.

272 Figure 7 contains a graphical representation of these results.

273 **Disease-related health episodes and resource use**

274 Select disease-related health episode and resource use data were also collected at the
275 beginning of the study and captured after 90 days of treatment with each device.
276 Baseline average number of episodes in the 12 months before study enrollment was
277 0.6 ± 0.1 for cellulitis and 0.3 ± 0.1 for ulceration. For resource use, the baseline
278 average number of days in the 12 months before study enrollment for hospitalization
279 associated with complications from lymphedema was 1.0 ± 0.4 , and for use of
280 compression stockings, it was 304.3 ± 14.6 . The average number of lymphedema-
281 related physical therapy visits in the 12 months before study was 19.5 ± 3.7 .

282 During the NPCD treatment period, for both the PLED and Non-PLED subsets, no
283 episodes of cellulitis, ulceration, or hospitalization were reported. Average lymphedema-
284 related physical therapy visits during this 90-day study period were found to be 0 visits
285 for the PLED subset and 0.39 visits for the Non-PLED subset, all for the NPCD group.

286 During the APCD treatment period, there were a total of three cases (~4%) of cellulitis
287 reported and one case of ulceration reported (~1%), all of which were resolved with
288 medical intervention. The three cases of cellulitis were observed in the Non-PLED
289 subset, while the ulceration was observed in the PLED subset. A total of 8
290 hospitalization days were also reported during APCD treatment period. Four of these
291 were observed in the Non-PLED subset, while the other four were observed in the LED
292 subset. Average lymphedema-related physical therapy visits during this 90-day study
293 period were found to be 3.35 visits for the LED subset and 1.86 visits for the Non-PLED
294 subset, all for the APCD group.

295 A summary of the health resource utilization data is shown in Table 3a (PLED) and 3b
296 (Non-PLED).

297 **Discussion**

298 In the TEAYS study¹¹, a total of 71 patients (108 affected limbs) with lower extremity
299 lymphedema were analyzed. Compared with the APCD, the NPCD was associated with
300 a greater mean decrease in limb edema volume (a mean limb volume decrease of
301 369.9 ± 68.19 mL vs 83.1 ± 67.99 mL). Significant improvement in Quality of Life was
302 achieved for NPCD and but not for APCD treatment (score improvement of 1.01 ± 0.23
303 for NPCD vs 0.17 ± 0.18 for APCD). Patients reported greater adherence (81% vs
304 56%; $p \leq .001$) and satisfaction with the NPCD (78% vs 22%) compared with APCD.

305 In the current study, analysis of subset of 35 patients diagnosed with PLED were
306 performed. PLED patients achieved statistically greater mean limb volume reduction
307 ($424.4 \pm 100.9 \text{ mL}$) while on NPCD vs ($50.8 \pm 112.1 \text{ mL}$) for APCD. NPCD also achieved
308 significantly better improvement in overall Quality of Life (1.39 ± 0.39) vs. APCD
309 (0.18 ± 0.29). Statistically significant improvement in adherence was also observed while
310 on NPCD 81% vs APCD 49%. No device-related adverse events were reported.

311 Specifically looking at the mean edema reduction, PLED patients treated with NPCD
312 demonstrated greater benefit than their Non-LED counterparts. For those PLED
313 patients, the mean limb volume reduction achieved was 424.4 mL while treated with
314 NPCD compared with 50.8 mL treated with APCD. That -is, those PLED patients
315 appeared to have achieved seven times greater edema reduction while on NPCD
316 treatment arm. In contrast, non-LED patients also achieved better mean limb volume
317 reduction while on NPCD compared to APCD, the difference, while statistically
318 significant are not as pronounced 317.3 mL for NPCD and 114.9 mL for APCD.
319 Furthermore, the PLED cohort appeared to achieve more reduction in both the midfoot
320 and metatarsal foot regions while on NPCD (0.29 and 0.43 vs APCD 0.17 and 0.23
321 respectively) numerically favoring NPCD.

322

323 A 3-year retrospective analysis involving 440 eligible patients with lower extremity
324 lymphedema showed that CVI (phlebolymphe dema), not cancer, was the predominant
325 cause of lower extremity lymphedema¹². These finding supports the well accepted
326 understanding that phlebolymphe dema (LED) stems from chronic venous insufficiency
327 (CVI), where damaged venous valves or alternatively venous occlusion from prior
328 injury/thrombosis impair venous return, and eventually contribute to lymphatic fluid
329 overload. As highlighted in the TEAYS study, NPCD differs from pneumatic
330 compression in both mechanism and design. Pneumatic compression provides
331 sequential compression while requiring patient to be immobilized and remain in a supine
332 position for the duration of the treatment. In a stark contrast, NPCD utilizes both static
333 and sequential gradient compression as well as provide added benefits for ambulation
334 and subsequent activation of calf and thigh pumps¹³—both of which critical to managing
335 phlebolymphe dema and venous health. This mobility-driven approach reflects the
336 contemporary approach to managing LED acknowledges the lymphatic-venous¹⁴⁻¹⁶
337 connection in fluid management.

338 The TEAYS study and this subsequent LED sub analysis have shown that NPCD not
339 only enhances treatment adherence but also provide additional plausible explanations
340 for combined effects of synergistic modalities that have been well established in prior
341 published literature and consensus;¹⁷ supporting a comprehensive treatment
342 lymphatic drainage and venous return.

343 **Limitations**

344 Common biases in crossover studies include unintended biases such as order of which
 345 device was first used and whether there is an unforeseen carryover effect from one
 346 treatment arm to the comparator. Additional analyses were performed to examine effect
 347 with respect to order of device. Additionally, despite the initial 30-day washout period,
 348 some patients who have prior experience with an APCD may retain preconceived
 349 notions about that device, which may have influenced compliance.

350 Conclusion

351 CVI associated with lymphatic damage (phlebolymphe^dema) is a predominant cause of
 352 lower extremity lymphedema arises from a compromised or poorly functioning venous
 353 system. While traditional recommendations like compression and elevation remain
 354 foundational, a significant treatment gap exists that addresses not only objective clinical
 355 outcomes, such as limb volume reduction, but also factors like quality of life (QoL) and
 356 patients' adherence to daily treatment regimens. The TEAYS study demonstrated that
 357 NPCD is a clinically effective and safe treatment option for lower extremity lymphedema
 358 while demonstrating superiority in key outcomes such as limb volume reduction, quality
 359 of life, and adherence compared to the traditional use of APCDs. The PLED subgroup
 360 analysis further reinforces these conclusions, showing that both PLED and non-LED
 361 patients gained meaningful clinical benefits from NPCD; with PLED patients potentially
 362 benefiting even more due to the comprehensive treatment approach offered by NPCD
 363 and the distinct pathophysiology involving CVI as the underlying etiology for PLED.

364 References

- 365
- 366 1. Rockson SG. Lymphedema. *Am J Med.* 2001;110(4):288-95.
 - 367 2. Farrow W. Phlebolymphe^dema — a common underdiagnosed and undertreated problem in the
 368 wound care clinic. *J Am Col Certif Wound Spec.* 2010;2(1):14-23.
 - 369 3. Lee BB. Phlebolymphe^dema: neglected outcome of combined venous and lymphatic insufficiency.
 370 *Vasc Spec Int.* 2020;36(1):1-3. doi:10.5758/vsi.2020.36.1.1
 - 371 4. Sleigh BC, Manna B. Lymphedema [Internet]. Treasure Island (FL): StatPearls Publishing; 2024 Jan-
 372 [updated 2023 Apr 19; cited 2025 Oct 3]. Available from:
 373 <https://www.ncbi.nlm.nih.gov/books/NBK537239/>
 - 374 5. Singh A, Zahra F. Chronic venous insufficiency [Internet]. Treasure Island (FL): StatPearls Publishing;
 375 2024 Jan- [updated 2023 Apr 27; cited 2025 Oct 3]. Available from:
 376 <https://www.ncbi.nlm.nih.gov/books/NBK587341/>
 - 377 6. Dean SM, Valenti E, Hock K, Leffler J, Compston A, Abraham WT. Clinical characteristics of lower
 378 extremity lymphedema in 440 patients. *J Vasc Surg Venous Lymphat Disord.* 2020;8(5):851-9.
 379 doi:10.1016/j.jvsv.2019.11.014
 - 380 7. Ulloa J, Cifuentes S. Phlebolymphe^dema: an up-to-date review. *Phlebolymphe^dology.* 2021;28(2):49-
 381 60.
 - 382 8. Rockson SG, Norman SA, Stout NL, Schmitz KH, Shaitelman SF, Armer JM, et al. Safety and
 383 effectiveness of a novel non-pneumatic active compression device for treating breast cancer-related

- 384 lymphedema: a multicenter randomized crossover trial (NILE). *J Vasc Surg Venous Lymphat Disord.*
385 2022;10(6):1359-66.e1.
- 386 9. Rockson SG, Norman SA, Stout NL, Schmitz KH, Shaitelman SF, Armer JM, et al. Clinical evaluation of
387 a novel wearable compression technology in the treatment of lymphedema: an open-label
388 controlled study. *Lymphat Res Biol.* 2022;20(2):125-32.
- 389 10. Rockson SG, Norman SA, Stout NL, Schmitz KH, Shaitelman SF, Armer JM, et al. A non-randomized,
390 open-label study of the safety and effectiveness of a novel non-pneumatic compression device
391 (NPCD) for lower limb lymphedema. *Sci Rep.* 2022;12:140005.
- 392 11. Barfield M, Winokur R, Berland T, Davis S, Ralph V, Chatham N, et al. Results from a comparative
393 study to evaluate the treatment effectiveness of a non-pneumatic compression device vs an
394 advanced pneumatic compression device for lower extremity lymphedema swelling (TEAYS study). *J*
395 *Vasc Surg Venous Lymphat Disord.* 2025;13(1):101965. doi:10.1016/j.jvsv.2024.101965
- 396 12. Dean SM, Valenti E, Hock K, Leffler J, Compston A, Abraham WT. Clinical characteristics of lower
397 extremity lymphedema in 440 patients. *J Vasc Surg Venous Lymphat Disord.* 2020;8(5):851-9.
- 398 13. Tauraginskii RA, Ivanov AV, Petrov DS, Smirnov IN, Volkov AA, Kuznetsov VV, et al. The human lower
399 leg muscle pump functions as a flow diverter pump, maintaining low ambulatory venous pressures
400 during locomotion. *J Vasc Surg Venous Lymphat Disord.* 2024;13(1):101996.
401 doi:10.1016/j.jvsv.2024.101996
- 402 14. Levick JR. Revision of the Starling principle: new views of tissue fluid balance. *J Physiol.* 2004;557(Pt
403 3):704-6. Available from: <https://pmc.ncbi.nlm.nih.gov/articles/PMC1665155/>
- 404 15. Woodcock TE, Woodcock TM. Revised Starling equation and the glycocalyx model of transvascular
405 fluid exchange: an improved paradigm for prescribing intravenous fluid therapy. *Br J Anaesth.*
406 2012;108(3):384-94.
- 407 16. Mortimer PS, Rockson SG. New developments in clinical aspects of lymphedema disease. *J Clin*
408 *Invest.* 2014;124(3):915-21.
- 409 17. Gloviczki P, Lawrence PF, Dalsing MC, Eklöf BG, Elias S, Gillespie DL, et al. The 2022 Society for
410 Vascular Surgery, American Venous Forum, and American Vein and Lymphatic Society clinical
411 practice guidelines for the management of varicose veins of the lower extremities. Part I: duplex
412 scanning and treatment of superficial truncal reflux. *J Vasc Surg Venous Lymphat Disord.*
413 2023;12(2):645-65. Available from: [https://www.jvsvenous.org/article/S2213-333X\(22\)00417-](https://www.jvsvenous.org/article/S2213-333X(22)00417-6/fulltext)
414 [6/fulltext](https://www.jvsvenous.org/article/S2213-333X(22)00417-6/fulltext)

415

Table 1: Study Demographics- Phlebolymphe^dema (PLED) vs Non-Phlebolymphe^dema (Non-PLED) Subsets

Number of Patients by Subtype	N=35 PLED	N=36 Non-PLED
Mean Age	58.5 ± 2.63 yrs*	59 ± 2.37 yrs*
Gender: M (F)	13 (22)	6 (30)
Race/Ethnicity		
Asian	0	2
Caucasian	31	28
African American	4	4
Hispanic	0	2
Average BMI	36.2 ± 1.68*	29.0 ± 1.24*
CEAP score (C1, C2, C3, C4, C5)	0, 10, 13, 5, 7	-
Affected Limbs: Unilateral (L/R) /Bilateral	16 (9/7) / 19	18 (9/9) / 18
Lymphedema history (years since diagnosis)	7.9 ± 1.2*	8.4 ± 1.42
Lymphedema Clinical Stage I, II, III	6, 20, 9	7, 24, 5
Percentage of subjects with sleep apnea	43%	25%

*Reported as Mean ± Standard Error

Table 2. Summary of data for PLED and Non-LED subset for the NPCD and APCD groups

	Device	PLED	Non-LED
Mean limb volume change (in mL)	NPCD	Mean with SE: 424.4 ± 100.9 Median: 349 <i>p</i> = .0011	Mean with SE: 317.3 ± 92.42 Median: 210 <i>p</i> = .0049
	APCD	Mean with SE: 50.8 ± 112.06 Median: 7.5 <i>p</i> = .65	Mean with SE: 114.9 ± 78.70 Median: 73 <i>p</i> = .42
	NPCD vs APCD	<i>p</i> = .0085	<i>p</i> = .034
Change in Foot Region (Mid foot) in CM	NPCD	Mean with SE: 0.29 ± 0.22 Median: 0.2 <i>p</i> = .18	Mean with SE: 0.20 ± 0.20 Median: 0.20 <i>p</i> = .32
	APCD	Mean with SE: 0.17 ± 0.24 Median: -0.1 <i>p</i> = .53	Mean with SE: -0.01 ± 0.17 Median: 0.1 <i>p</i> = .96
	NPCD vs APCD	<i>p</i> = .40	<i>p</i> = .14
Change in Foot Region (Meta-tarsal) in CM	NPCD	Mean with SE: 0.43 ± 0.17 Median: 0.6 <i>p</i> = .036	Mean with SE: 0.17 ± 0.17 Median: 0.20 <i>p</i> = .34
	APCD	Mean with SE: 0.23 ± 0.23 Median: 0.3 <i>p</i> = .34	Mean with SE: -0.17 ± 0.12 Median: 0.2 <i>p</i> = .13
	NPCD vs APCD	<i>p</i> = .36	<i>p</i> = .39
Overall LYMQOL	NPCD	Mean with SE: 1.39 ± 0.39 Median: 1.00 <i>p</i> = .015	Mean with SE: 0.64 ± 0.24 Median: 0 <i>p</i> = .103
	APCD	Mean with SE: 0.18 ± 0.29 Median: 0 <i>p</i> = .53	Mean with SE: 0.16 ± 0.23 Median: 0 <i>p</i> = .51
	NPCD vs APCD	<i>p</i> = .01	<i>p</i> = .10
LYMQOL – Function	NPCD	Mean with SE: 0.37 ± 0.12	Mean with SE: 0.11 ± 0.08
	APCD	Mean with SE: -0.02 ± 0.11	Mean with SE: 0.16 ± 0.08
	NPCD vs APCD	<i>p</i> = .008	<i>p</i> = .36
LYMQOL – Appearance	NPCD	Mean with SE: 0.34 ± 0.10	Mean with SE: 0.22 ± 0.08
	APCD	Mean with SE: 0.09 ± 0.08	Mean with SE: 0.12 ± 0.06
	NPCD vs APCD	<i>p</i> = .021	<i>p</i> = .15
LYMQOL – Symptom	NPCD	Mean with SE: 0.19 ± 0.09	Mean with SE: 0.12 ± 0.08
	APCD	Mean with SE: 0.02 ± 0.09	Mean with SE: 0.05 ± 0.06
	NPCD vs APCD	<i>p</i> = .05	<i>p</i> = .25
LYMQOL – Mood	NPCD	Mean with SE: 0.16 ± 0.13	Mean with SE: 0.10 ± 0.08
	APCD	Mean with SE: 0.01 ± 0.07	Mean with SE: 0.06 ± 0.05
	NPCD vs APCD	<i>p</i> = .11	<i>p</i> = .36
Adherence (in %)	NPCD	Mean with SE: 81 ± 4 Median: 92	Mean with SE: 80 ± 4.2 Median: 89
	APCD	Mean with SE: 49 ± 6.5 Median: 46	Mean with SE: 62 ± 5.3 Median: 69
	NPCD vs APCD	<i>p</i> ≤ .001	<i>p</i> = .0024

Table 3a. Health Resource Utilization, during the study with each device (PLED)

Subjects	With NPCD for 3 Months	With APCD for 3 Months
Phlebo subset who completed the study	35	35
Average cellulitis episodes	0	0
Average ulceration episodes	0	0.03 ± 0.03*
Average hospitalization days	0	0.12 ± 0.08*
Average # of physical therapy visits	0	3.35 ± 1.14*
Percentage of subjects reporting decreased use of compression garments	68%	12%

*Reported as Mean ± Standard Error

Table 3b. Health Resource Utilization, during the study with each device (Non-PLED)

Subjects	With NPCD for 3 Months	With APCD for 3 Months
Non-Phlebo subset who completed the study	36	36
Average cellulitis episodes	0	0.08 ± 0.05*
Average ulceration episodes	0	0
Average hospitalization days	0	0.11 ± 0.09*
Average # of physical therapy visits	0.39 ± 0.33*	1.86 ± 0.59*
Percentage of subjects reporting decreased use of compression garments	64%	6%

*Reported as Mean ± Standard Error




Figure 1. Study Design




Figure 2. Change in Mean Limb Volume Compared to Baseline, (a) PLED vs. (b) Non-LED Subsets




Figure 3. Mean Change in the Foot Region, (a) PLED vs. (b) Non-LED Subsets




Figure 4. Overall LYMQOL, (a) PLED vs. (b) Non-LED Subsets




Figure 5. LYMQOL Improvement, PLED vs. Non-PLED Subsets




Figure 6. Adherence to Treatment, PLED vs. Non-PLED Subsets




Figure 7. Subject Preference Questionnaire, PLED vs. Non-PLED Subsets